



OFFICE OF THE DISTRICT HEALTH OFFICER
GULKADA DISTRICT SWAT Phone No: 0946-9240139, Fax No: 0946-9240215
Email: edohswat@yahoo.com

TENDER NOTICE FOR THE REPAIR OF MACHINERY & EQUIPMENT.

Sealed & postal Tender in accordance with KPPRA rules—single stage two envelope procedure are invited from the reputed firms/ companies—GST registered, NTN holders, on ATL, having registration with Pakistan Engineering Council for relevant category, financially sound, have experience in the repair of Machinery and Equipment for the repair of Hospitals equipment along with provision of spare parts and related accessories, under the District Govt; allocated budget during financial year 2016-17.

The quotation should reach office of the undersigned up to 10/05/2017 at 11:00 AM with 2% bid security which will be opened on the same day in the presence of representative of firms if they want and in case of the leave, the next day.

The undersigned can reject any quotation as per rules.

Terms, Conditions and detail of repair work and specification can be obtained from Accounts Section of DHO Swat office during working hours and can also be downloaded from <http://www.healthkp.gov.pk> website.


District Health Officer
Swat at Gulkada.

Repair file.



OFFICE OF THE DISTRICT HEALTH OFFICER
GULKADA DISTRICT SWAT Phone No: 0946-9240139, Fax No: 0946-9240215
Email: edohswat@yahoo.com

TENDER NOTICE FOR THE REPAIR OF MACHINERY & EQUIPMENT.

Sealed/postal Tender in accordance with KPPRA rules—single stage two envelope procedure are invited from the reputed firms/ companies—GST registered, NTN holders, on ATL, having registration with Pakistan Engineering Council for relevant category, financially sound, have experience in the repair of Machinery and Equipment under the District Govt ;allocated budget for the repair of Hospitals equipment along with provision of spare parts and related accessories during financial year 2016-2017.

The quotation should reach office of the undersigned up to 10/05/2017 at 11:00 AM with 2% bid security.

Terms, Conditions and detail of repair work and specification can be obtained from DHO Swat office during working hours and can also be downloaded from <http://www.healthkp.gov.pk> website.

Terms & Condition:-

- A) Firms shall have experience and have their complete set up for repair/ maintenance of all or any specific item/equipment mentioned.
- B) Firms must have General Sales Tax registration, vendor number from AG office Peshawar and National Tax number duly attested.
- C) Tender must be accompanied by a 2% CDR in shape of bank draft/pay order in favour of District Health Officer Swat as earnest money, without which the offer will be rejected.
- D) The earnest money of unsuccessful bidders will be refunded on finalization of the tender and in the case of successful bidder it will be retained till 30.06.2017.
- E) Firms shall be bound to provide/install genuine parts in each items/equipment. In case, the spare parts replaced are found sub-standard or second hand, the firms will be blacklisted. And the security amount/money will be forfeited.
- F) Firms shall be bound to return the replaced/old spare item replaced or repaired.
- G) The intending firms must provide a list of its relevant experience and past performance capabilities with respect of personnel, equipment and plant.
- H) The firm will provide an affidavit on Rs.100/- stamp papers duly attested by the oath commissioner to the effect that it has never been blacklisted.
- I) If parts is essential for normal function of the equipment/machinery but is omitted/not included in the repair list then that part should be included in the Tender/Bid together with its price.
- J) Firm will be responsible for normal functionalization of the machinery /equipment.
- K) The successful bidder will execute an agreement with District Health Officer Swat at Gulkada on Rs.100/- stamp paper duly attested by Oath commissioner that the firm will finish the work within the stipulated time and without cost escalation
- L) The DHO Swat reserves the right to accept or reject any or all bids/proposals as per KPPRA rules.
- M) The bids/proposals should reach the office of the District Health Officer Swat at Gulkada on or before 10/05/2017 at 11:00 AM. The bids will be opened on the same day at 12:00 PM in the presence of the bidders or their representatives.
- N) 10% Bid security will be deducted and will be released after warranty period (one year warranty).
- O) Taxes if any will be deducted as per rules.
- ~~P) 10% Bid security will be deducted and will be released after warranty period (one year warranty).~~
- ~~Q) Taxes if any will be deducted as per rules.~~

District Health Officer
Swat at Gulkada



OFFICE OF THE DISTRICT HEALTH OFFICER
GULKADA DISTRICT SWAT Phone No: 0946-9240139, Fax No: 0946-9240215
 Email: edohswat@yahoo.com

Detail of repair work, parts, specification, Machinery & Equipment in various Hospitals in District Swat.

S. No.	Hospital Name	Equipment Name	Model/Serial No.	Made/ manufacture r	Remarks/parts component required
1.	CH Barikot	X-Ray M/C 100 MA		Siemen Germany	Tube Head need repair / replacement lying at Batkhela workshop . specification of Tube Head-100 mA.
	CH Barikot	Chargeable X-Ray M/C 500 MA			HT- transformer -500 mA is out of order, battery bank.
2.	RHC Chuprial	Dental Unit	EXCEL	Pak Made	i. Complete over hauling. ii. Internal water/Air tubing iii. Valves complete. iv. Chair up/Down system. V. Light system & bulb iv) Hand pieces. High speed and slow speed.
3.	THQ Hospital Matta	Baby warmer 02 Nos. Infant warmer 02 No.	Type-B/1035552,69. Type-B/ HK93/23071103003/	Americ e serve. Ningbo David China.	Needs Transformers 02 Nos. Needs Transformers 02 Nos.
4.	THQ:Hospital Khwazakhela.	ECG Cable 2- No.	Cardio care 2000 EKG 2000	Bionet.	Needs replacement.
5.	THQ Hospital Matta	Infant Incubator 04 No's	C2HS-IC/03 YP200/06D70202 YP-90AB/08100302- F-2855CB/8807.0042	USA.-55 China.-500 China China	Require repair of i. Main PCB ii. Humidity trie missing i. Main PCB ii. Sensor. i. control missing. i. Control PCB II. Panel Board.
6.	CH Madyan	Repair of AVR for X- Ray M/C 300 MA.	DH325P/960277	Donga Korea	The M/C being recently shifted from CH Kalam & Installed by private technician but leaking the function of Bucky system. Also, the AVR provide is unable to compensate for the low voltage & herby M/C shows overloa appreciably low. repair of AVR with X Ray plant.
7.	CH Kalam	Dental Plant			Compressor & UP Down not working repair and Complete overhauling.
8.	CH Kabal	X-Ray 300 MA Portable	Model DMHC 325 S/No.0073	Listen Korea.	HT Tank leaks. HT control board.
9.	THQ Hospital Khwazakhela.	1- Dental Unit	Model Indus S/No.119511102	PAK Made	-position switch. -Solenoid valve. -Pushing.